Micropipette Justers



Description:

The Micropipette Adjuster series is designed for precise and reliable liquid handling across a wide range of laboratory applications. Available in single-channel models covering volumes from 0.1 µl up to 10 ml, these micropipettes deliver excellent accuracy and reproducibility for routine and advanced workflows. Featuring fine volume increments, ergonomic handling, and consistent performance, the Micropipette Adjuster ensures user comfort during repetitive pipetting while maintaining high measurement precision. Ideal for molecular biology, microbiology, clinical, and research laboratories, this range offers dependable results for ultra-low to high-volume dispensing tasks.

FDA APPROVED
CE
ISO 16602

Micropipette Justers

Volume Range	Channel	Increment	Measuring Volume	Inaccuracy	Imprecision
0.1–2.5 µl	1-ch	0.05 µl	2.5 µl	±2.50%	2.00%
			1.25 µl	±3.00%	3.00%
			0.25 µl	±12.00%	6.00%
0.5–10 µl	1-ch	0.10 µl	10 µl	±1.00%	0.80%
			5 µl	±1.50%	1.50%
			1 µl	±2.50%	1.50%
2–20 µl	1-ch	0.50 µl	20 µl	±0.90%	0.40%
			10 µl	±1.20%	1.00%
			2 µl	±3.00%	2.00%
5–50 µl	1-ch	0.50 µl	50 µl	±0.60%	0.30%
			25 µl	±0.90%	0.60%
			5 µl	±2.00%	2.00%
10–100 µl	1-ch	1.00 µl	100 µl	±0.80%	0.20%
			50 µl	±1.00%	0.40%
			10 µl	±3.00%	1.00%
20–200 µl	1-ch	1.00 µl	200 µl	±0.60%	0.20%
			100 µl	±0.80%	0.30%
			20 µl	±0.60%	0.80%
100–1000 µl	1-ch	5.00 µl	1000 µl	±0.70%	0.20%
			500 µl	±2.00%	0.25%
			100 µl	±0.50%	0.70%
1–5 ml	1-ch	50 µl	5000 µl	±0.60%	0.20%
			2500 µl	±0.70%	0.30%
			1000 µl	±2.50%	0.30%
1–10 ml	1-ch	100 µl	10000 µl	±0.60%	0.20%
			5000 µl	±1.20%	0.30%
			1000 µl	±3.00%	0.60%

FDA APPROVED
CE
QUALITY ASSURED FIRM ISO 16602